

Agence Nord: Campus Effiscience 2 Rue Jean Perrin - Bât D 14460 Colombelles tél: 02.31.34.50.74 fax: 02.31.34.55.17



Parc Club des Tanneries 2 Rue de la Faisanderie 67380 Lingolsheim tél: 03.88.04.01.81 fax: 03.68.93.01.52

Agence Est:

www.deltalabo.fr

Agence Sud: Bât Le Venango, 392

Bât Le Venango. 392 Rue Jean Dausset AGROPARC - BP11575 84916 Avignon Cédex 9 tél : 04.90.27.17.95 fax : 04.90.27.17.52

Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412



The Compact and Efficient Solution

- 4 independently controlled test stations
- Multiple operator safety features
- Motorised raising and lowering
- Automated test sequence
- Intermediate scraping position
- Integrated timing
- Easy handling and removal of sample
- No removal of the paddles required
- Guaranteed paddle rotation speed
- Non-reflective enhanced LED lighting
- Large viewing window
- Large LCD touchscreen display
- Small footprint
- Intuitive software package

Petroleum Oils • Synthetic Fluids



Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412

Innovative Design

The Seta Herschel Emulsifier is a compact and efficient benchtop instrument designed to measure the ability of petroleum oils and synthetic fluids to separate from water. The instrument incorporates 4 test stirrers, with independent control and motorised raising and lowering. Samples can be tested simultaneously or individually to suit laboratory requirements.

Each stirrer is held by a self centering collet to ensure concentricity within 1mm and can remain in situ when removing or replacing cylinders. Positioning within 6mm from the bottom of the cylinder is automatically achieved using a motorised actuator.

After the stirring period the motorised head raises the stirrer until clear of the graduate cylinder to allow for wiping and on completion of the test, the stirrer will revert to the higher, home position.

Test cylinders are located in a temperature controlled bath with an adjustable set point of either 54 or 82°C. A large viewing window and non-reflective LED lighting assists operator reading. The bath top plate is designed to allow cylinders to tilt for easy removal. A drain valve is provided for service and maintenance.

A large colour touch screen display is used to initiate tests and provides automated sequencing with an audible and visual reminder at each recording interval. Custom test parameters are also user adjustable.




Typical Applications

• Petroleum Oils, Synthetic Fluids

Test Method

ASTM D1401. A sample is mixed with water, heated in the bath and stirred for 5 minutes whilst immersed. After stirring, the time taken for the resultant emulsion to separate is recorded every 5 minutes until the emulsion separates completely or reduces to 3 ml or less.

Operator Interface and Integrated Timing



Independent control allowing for simultaneous or individual testing



Timer 'prompt' at recording intervals to aid visual measurements



User adjustable test parameters


For more information please visit: www.stanhope-seta.co.uk



Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412



Operator Safety

- Obstruction override if paddles are placed under any resistant load
- Duplex protective viewing glass
- Automatic high temperature cut out
- Overflow prevention
- Automatic low-level liquid trip
- Abort button for emergency all stop


Ease Of Use

- Fully automated paddle movement with motorised raising and lowering
- Easy sample handling without removing the stirrer paddle
- Self-centering collet to ensure concentricity within 1mm
- Positioning within 6mm from the bottom of the cylinder automatically achieved
- Large LCD touch screen display
- Integrated timer with audible and visual reminders for recording intervals



Precision and Accuracy

- Non-reflective enhanced LED lighting for improved measurement precision
- User adjustable sample temperature stabilisation time
- Post stirring, intermediate scraping position ensures accurate readings
- Guaranteed paddle rotation speed, regardless of viscosity
- Large viewing window



Enhanced Test Throughput

- 4 independent test stations reduce waiting time
- Rapid bath heat up time



For more information please visit: www.stanhope-seta.co.uk



Herschel Emulsifier 96700-2

Water Separability of Petroleum Oils and Synthetic Oils

ASTM D1401; ISO 6614; FTM 791-3201; IP 412

Technical Specifications

| Herschel Emulsifier | |
|----------------------------|---|
| Bath volume | 5 litres |
| Bath liquid | Water or White oil |
| Sample size | 40 ml oil |
| | 40 ml distilled water |
| | 1% sodium chloride solution or synthetic seawater |
| Test temperatures | 54 °C and 82 °C |
| Bath temperature stability | ±1°C |
| Stirrer speed | 1500 ± 15 rpm |
| Voltage options | 110 Vac, 50/60 Hz |
| | 240 Vac, 50/60 Hz |
| Current | 10 A (max)* - 110 Vac model |
| | 5 A (max)* - 220 Vac model |
| Display | LCD touchscreen |
| Size: (H x W x D) | 890 x 450 x 450 mm |
| Weight (empty) | 49.5 kg |

^{*} In-rush current will be higher

Minimal Servicing

| Herschel Emulsifier | | |
|---------------------|--------------------------|--|
| Weekly | Tighten paddle collets | |
| Annually | Temperature calibration | |
| | Paddle speed calibration | |

The instrument is supplied with the following:

- 4 graduated glass cylinders
- Screen protector
- Spatula

Accessories

| Part Number | | Description |
|-------------|--|--|
| 51000-0 | | Digital thermometer 6 x 210 mm probe. 5 point UKAS calibration at -50, 0, 50, 100 and 150 °C |
| 94630-0 | The state of the s | Bath oil, blended with anti-oxidant Suitable for temperatures between 40 and 85 °C. Supplied in a 5 litre container |
| 99960-2 | 1000 EXCESS | Tachometer Accuracy ±1.75 rpm at 1500 rpm |
| 11227-0 | The state of the s | Synthetic seawater solution Supplied with Certificate of Conformance to ASTM D665 |

For more information please visit: t6: 02.31.34.50.74 fax: 02.31.34.55.17

Agence Nord:
Campus Effiscience
2 Rue Jean Perrin - Bât D
14460 Colombelles
tél : 02.31.34.50.74



Agence Est:
Parc Club des Tanneries
2 Rue de la Faisanderie
67380 Lingolsheim
tél: 03.88.04.01.81
fax: 03.68.93.01.52

Sud:
www.deltalabo.fr
arc - BP11575

info@deltalabo.fr

Agence Sud

Bât Le Venango. 392 Rue Jean Dausset AGROPARC - BP11575 84916 Avignon Cédex 9 tél : 04.90.27.17.95 fax : 04.90.27.17.52